Programmer's Manual

Digital Gamma Finder (DGF)

Model DGF-4C

Version 3.00, April 2003

X-Ray Instrumentation Associates

8450 Central Ave
Newark, CA 94560 USA

Phone: (510) 494-9020; Fax: (510) 494-9040
http://www.xia.com

Disclaimer

Information furnished by XIA is believed to be accurate and reliable. However, XIA assumes no responsibility for its use, or for any infringement of patents, or other rights of third parties, which may result from its use. No license is granted by implication or otherwise under the patent rights of XIA. XIA reserves the right to change the DGF product, its documentation, and the supporting software without prior notice.

1 Overview 1
2 DGF-4C C Library 2
2.1 C_Dgf4c_Hand_Down_Names 2
2.2 C_Dgf4c_Boot_System 3
2.3 C_Dgf4c_User_Par_IO 3
2.4 C_Dgf4c_Acquire_Data 4
2.5 C_Dgf4c_Set_Current_ModChan 5
2.6 C_Dgf4c_Buffer_IO 6
2.7 DGF-4C C Library Compile Options 6
3 Control DGF-4C Modules via CAMAC 8
3.1 Initializing 8
3.1.1 Initialize Global Variables 8
3.1.2 Boot DGF4C Modules 9
3.2 Setting DSP variables 10
3.3 Access spectrum memory or list mode data 10
3.3.1 Access spectrum memory 10
3.3.2 Access list mode data 11
4 User Accessible Variables 13
4.1 Module parameters 13
4.2 Channel variables 18
4.3 ADC data 28
5 Control Tasks 30
6 Appendix A — User supplied DSP code 34
6.1 Introduction 34
6.2 The development environment 34
6.3 Interfacing user code to XIA's DSP code 34
6.4 The interface 35
6.5 Debugging tools 38

1 Overview

This manual is divided into three major sections. The first section is a description of the DGF-4C C Library which is currently used by the DGF-4C Viewer. Advanced users can build their own user interface using the user accessible functions in the library. The second section is a reference guide to program the DGF-4C modules via CAMAC. This will be interesting to those users who want to integrate the DGF-4C modules into their own data acquisition system. The third section describes those user accessible variables that control the functions of the DGF-4C modules. Those advanced and curious users can user this section to better understand the operation of the DGF-4C. Additionally, this manual also includes instructions on how to write User DSP code.

The scope of this document is all DGF-4C modules with serial numbers D1100 through D1199 and E1200 through E1299. Modules with serial numbers E1200 through E1299 have 14-bit ADCs as opposed to the 12-bit ADCs on other modules. Modules with serial numbers D1100 through D1199 and E1200 through E1299 have 4k FIFOs, which allow storing up to 4096 ADC samples for each channel.

2 DGF-4C C Library

The DGF-4C C Library contains a group of C functions used to control the DGF-4C modules. It can be compiled as either a XOP file currently used by the DGF-4C Viewer, or a dynamic link library (DLL) or static library used by customized user interface or applications. The library provides users six functions which can be used to fully control the modules. These six functions were described in sections 2.1 to 2.6. Section 2.7 gives the different options of compiling the DGF-4C C Library.

2.1 C_Dgf4c_Hand_Down_Names

Function prototype

long C_Dgf4c_Hand_Down_Names(char *Names[], char *Name)

Parameter description

Name: A string variable used to direct which set of names to be handed down. It can be one of the following four choices: “ALL_FILES”, “MODULE_GLOBAL_NAMES”, “GLOBAL_DATA_NAMES”, or “USER_VAR_NAMES”.

Names: A two dimensional string array containing either the file names or the global parameter names. It can be one of the following four sets of names corresponding to the other parameter Name:

All_Files is a string array which has (MAX_NUMBER_OF_MODULES+5) elements. Currently MAX_NUMBER_OF_MODULES is defined as 24. The first five elements of All_Files are the name of system FPGA file, DSP code binary file, DSP I/O parameter values file, DSP code I/O variable names file, and DSP code all variable names file. The remaining elements are FIPPI file names for each module. All file names should contain the complete path name. Note: all modules use the same system FPGA and DSP codes, but could use different FIPPI files.

Module_Global_Names is a string array containing global variable names which are applicable to all modules, e.g. number of modules in the crate, the CAMAC controller type, the SCSI number, and the CAMAC crate ID, etc. Module_Global_Names currently can hold 64 names. If less than 64 names are needed (which is the current case), the remaining names should be defined as empty strings.

Global_Data_Names is a string array containing global variable names which are applicable to each individual module, e.g. module number, module CSR, coincidence pattern, and run type, etc. Global_Data_Names currently can hold 64 names. If less than 64 names are needed (which is the current case), the remaining names should be defined as empty strings.

User_Var_Names is a string array containing variable names which are applicable to individual channels of individual modules, e.g. channel CSR, filter rise time, filter flat top, gain, and offset, etc. User _Var_Names currently can hold 64 names. If less than 64 names

are needed (which is the current case), the remaining names should be defined as empty strings.

Function description

This function can be used to download system file names or user and global names from the host to the DGF-4C C Library.

2.2 C_Dgf4c_Boot_System

Function prototype

long C_Dgf4c_Boot_System(long Boot_Pattern)

Parameter description

Boot_Pattern is a bit mask:
- Bit 0: Boot system FPGA
- Bit 1: Boot FIPPI
- Bit 2: Boot DSP
- Bit 3: Load DSP parameters
- Bit 4: Apply DSP parameters (call Set_DACs and Program_FIPPI)

Under most of the circumstances, all the above tasks should be executed to initialize the DGF-4C modules, i.e. the Boot_Pattern should be 0x1F.

Function description

This function can be used to boot all DGF-4C modules in the system. Before booting the modules, this function initializes the CAMAC communication port and if CAMAC Master is going to be used, loads the CAMAC station number register.

2.3 C_Dgf4c_User_Par_IO

Function prototype

long C_Dgf4c_User_Par_IO(double *User_Par_Values, char *User_Par_Name, long direction)

Parameter description

direction = 0: download (write) parameter values from the host to the library;
1: upload (read) parameter values from the library to the host.

User_Par_Name is a string variable which directs what type of parameter values to be transferred. User_Par_Name can be one of the following four choices:

"MODULE_GLOBAL_NAMES" – to transfer global variable values applicable to all modules.

"GLOBAL_DATA_NAMES" – to transfer global variable values applicable to individual modules.

"SYNCH_WAIT" and "IN_SYNCH" – to broadcast SYNCH_WAIT or IN_SYNCH to all modules.

Name in array User_Var_Names (see Table 3)– to transfer variable values applicable to individual channels of individual modules.

User_Par_Values is a double precision array containing the parameter values to be transferred. For different User_Par_Name, different User_Par_Values array should be used. Totally three User_Par_Values arrays should be defined. All of these three arrays are one-dimensional arrays. The corresponding relationship between the User_Par_Values array and User_Par_Name is listed in Table 1.

Table 1: The Combination of User_Par_Name and User_Par_Values.

User_Par_Name	User_Par_Values		
	Name	Size	Data Type
MODULE_GLOBAL_NAMES	Module_Global_Values	64	Double precision
GLOBAL_DATA_NAMES	Global_Data_Values	64×24	Double precision
SYNCH_WAIT or IN_SYNCH	Global_Data_Values	64×24	Double precision
Name in array User_Var_Names	User_Values	64×24×4	Double precision

Function description

This function downloads or uploads global parameter or user values between the host and the library. For those DSP I/O parameters, this function calls other functions to convert these parameter values to DSP recognizable values or verse versa.

2.4 C_Dgf4c_Acquire_Data

Function prototype

long C_Dgf4c_Acquire_Data(long Run_Type, unsigned int *User_data, char *file_name)

Parameter description

User_data is an unsigned 32-bit integer array which holds data transferred between the modules and the host.

Run_Type is a 16-bit word whose lower 12-bit specifies either data_run_type or control_task_run_type and upper 4-bit specifies actions (start\stop\poll) as described below.

Lower 12-bit:

0x100,0x101,0x102,0x103	list mode runs
0x200,0x201,0x202,0x203	fast list mode runs
0x301	MCA run

	0x1 -> 0x15	control task runs
	0x3	adjust offsets
	0x4	acquire ADC traces
Upper 4-bit:		
	0x1000	start new run
	0x2000	resume run
	0x3000	stop run and automatically store spectrum data
	0x4000	poll
	0x500x	list mode special runs
	0x5000	Parse list mode data file
	0x5001	Locate list mode traces
	0x5002	Read list mode traces
	0x5003	Read list mode energies
	0x6000	stop list mode run during repeated runs
	0x7000	manually read spectrum from module
	0x8000	manually read spectrum from a MCA file

file_name is a string variable which specifies the name of the output file. file_name needs to have complete path.

Function description

This function acquires ADC traces, MCA spectrum, or list mode data. The string variable file_name needs to be specified when stopping a MCA run or list mode run in order to save the MCA spectrum data or list mode data into a file, or in those special list mode runs where the file_name should point to the saved list mode data file. In all other cases, file_name can be specified as an empty string. The unsigned 32-bit integer array User_data is used only for acquiring ADC traces (control task 0x4), reading out the list mode traces and energies in those special list mode runs, or reading out MCA spectrum. In all other cases, User_data can be any unsigned integer array with arbitrary size. Make sure that the User_data has the correct size and data type before reading out ADC traces, list mode traces or energies, or MCA spectrum.

2.5 C_Dgf4c_Set_Current_ModChan

Function prototype

long C_Dgf4c_Set_Current_ModChan(unsigned short Module, unsigned short Channel)

Parameter description

Module is an unsigned 16-bit integer which specifies the current module to be set. Module should be in the range of 1 to 23.

Channel is an unsigned 16-bit integer which specifies the current channel to be set. Channel should be in the range of 0 to 3.

Function description

This function sets the current module number and channel number.

2.6 C_Dgf4c_Buffer_IO

Function prototype

long C_Dgf4c_Buffer_IO(unsigned short *Values, unsigned short type, unsigned short direction, char *file_name)

Parameter description

The parameter Values is an unsigned 16-bit integer array used for data transfer between the host and the library. The parameter type specifies the I/O type. The parameter direction indicates the data flow direction. The string variable file_name contains the name of the settings file. Different combinations of the three parameters - Values, type, direction – designate different I/O operations as listed in Table 2.

Table 2: The Description of Function C_Dgf4c_Buffer_IO.

Type	Direction	Values	I/O Operation
0	0	DSP I/O variable values	Write DSP I/O variable values to modules
	1		Read DSP I/O variable values from modules
1	1	DSP all variable values	Read all DSP variable values from modules
2	0	N/A	Save current settings in all modules to a file
	1		Read settings from a file and apply to all modules
3	0	Values[0] – source module number; Values[1] – source channel number; Values[2] – copy/extract pattern bit mask; Values[3], Values[4], … - destination channel pattern	Extract settings from a file
	1		Copy settings from a source module to destination modules

Function description

This function downloads or uploads DSP settings between the host, the library and the DGF-4C modules; saves DSP settings into a settings file or loads DSP settings from a settings file and applies to all modules present in the system; or copies settings from one module to others or extracts settings from a settings file and applies to the modules present in the system

2.7 DGF-4C C Library Compile Options

DGF4C C Library can be compiled as either an Igor XOP used in the DGF-4C Viewer, or a standalone C-Library free of Igor stuff. The latter can be used by advanced users to develop their own data acquisition systems.

The following table lists the required files for these two options.

Table 3: Compilation Options of the DGF-4C C Library.

Compile Option	Required Files		
	C source files	C header files	Library files
Standalone C-Library	boot.c, camac.c, camacdll.c, CC32.c, Communication.c, dgf4c_c.c, utilities.c	boot.h, Camacdll.h, globals.h, sharedfiles.h, utilities.h, Libcc32.h, vpcic32d.h, Winaspi.h	pcicc32_ni.lib, pcicc32_ni.dll
Igor XOP	boot.c, camac.c, camacdll.c, CC32.c, Communication.c, dgf4c_c.c, utilities.c, dgf4c_iface.c, dgf4c_igor.c, Dgf4cWinCustom.rc	boot.h, Camacdll.h, globals.h, sharedfiles.h, utilities.h, Libcc32.h, vpcic32d.h, Winaspi.h, dgf4c_iface.h	pcicc32_ni.lib, pcicc32_ni.dll

The Igor XOP option also needs the following files in the Igor XOP Library provided by WaveMetrics.

IgorXOP.h, VCExtraIncludes.h, Xop.h, XOPResources.h, XOPStandardHeaders.h, XOPSupport.h, XOPSupportWin.h, XOPWinMacSupport.h, XOPSupport x86.lib, and IGOR.lib.

3 Control DGF-4C Modules via CAMAC

3.1 Initializing

The DGF-4C C Library has greatly simplified the initialization of DGF-4C modules via CAMAC compared to the previous implementation by XIA using a set of CAMAC commands. We describe below how to initialize DGF-4C modules in a CAMAC crate using the functions described in section 2. As an example, we assume two DGF-4C modules – one revision-D and one revision-E module – sit in slot 1 and 2, respectively. The CAMAC controller sits in slot 24.

3.1.1 Initialize Global Variables

As discussed in section 2.3, we assume that three global variable arrays have been defined: Module_Global_Values, Global_Data_Values and User_Values. For these three global variable arrays, we also need to define three global name arrays: Module_Global_Names, Global_Data_Names and User_Var_Names. Table 4 lists the names for these name arrays. The order of placing these names into the name array is not important since the C Library uses search functions to locate each name at run time.

Table 4: Contents of Global Name Arrays.

Array	Names
Module_Global_Names	NUMBER_MODULES, CONTROLLER_ID, SCSI_BUS, CRATE_ID, CAMAC_MASTER, FAST_CAMAC, LAM_ENABLE, C_LIBRARY_RELEASE, C_LIBRARY_BUILD, SLOT_WAVE
Global_Data_Names	MODULE_NUMBER, MODULE_CSRA, MODULE_CSRB, MODULE_FORMAT, MAX_EVENTS, COINCIDENCE_PATTERN, COINCIDENCE_WAIT, SYNCH_WAIT, IN_SYNCH, RUN_TYPE, BUFFER_HEAD_LENGTH, EVENT_HEAD_LENGTH, CHANNEL_HEAD_LENGTH, OUTPUT_BUFFER_LENGTH, NUMBER_EVENTS, RUN_TIME
User_Var_Names	CHANNEL_CSRA, CHANNEL_CSRB, ENERGY_RISETIME, ENERGY_FLATTOP, TRIGGER_RISETIME, TRIGGER_FLATTOP, TRIGGER_THRESHOLD, VGAIN, VOFFSET, TRACE_LENGTH, TRACE_DELAY, PSA_START, PSA_END, EMIN, BINFACTOR, TAU, BLCUT, XDT, CFD_THRESHOLD, MULTIPLICITY_PULSE_WIDTH, LIVE_TIME, INPUT_COUNT_RATE

Additionally, a string array All_Files containing the file names for the initialization is also needed. Table 5 lists the file names needed to initialize two DGF-4C modules.

Table 5: File Names in All_Files.

All_Files	File Name
All_Files[0]	C:\XIA\DGF4C\Firmware\dgf4c.bin
All_Files[1]	C:\XIA\DGF4C\DSP\DGFcodeE.bin
All_Files[2]	C:\XIA\DGF4C\Configuration\test.itx
All_Files[3]	C:\XIA\DGF4C\DSP\DGFcodeE.var
All_Files[4]	C:\XIA\DGF4C\DSP\DGFcodeE.lst
All_Files[5]	C:\XIA\DGF4C\Firmware\fdgf4c4d.bin
All_Files[6]	C:\XIA\DGF4C\Firmware\fdgf4c4e.bin

The global variable array, Module_Global_Values, also needs to be initialized before C Library functions can be called to start the initialization.

Table 6: Initialization of Module_Global_Values.

Module_Global_Names	Module_Global_Values
NUMBER_MODULES	2
CONTROLLER_ID	0 (0: J73A, 1: CC32, 2: offline)
SCSI_BUS	0
CRATE_ID	1
CAMAC_MASTER	1 (1: enable, 0: disable)
FAST_CAMAC	0 (1: enable, 0: disable)
LAM_ENABLE	0 (1: enable, 0: disable)
SLOT_WAVE[0]	24 (CAMAC MASTER CONTROLLER)
SLOT_WAVE[1]	1 (Module #1 sits in slot #1)
SLOT_WAVE[2]	2 (Module #2 sits in slot #2)

3.1.2 Boot DGF4C Modules

The boot procedure for DGF-4C modules includes the following steps. First, all the global parameter names should be downloaded by calling function C_Dgf4c_Hand_Down_Names. Then function C_Dgf4c_User_Par_IO should be called to initialize the global variable array Module_Global_Values. After that, function C_Dgf4c_Hand_Down_Names should be called to download the file name array All_Files. Finally, function C_Dgf4c_Boot_System should be called to boot the modules. The following code is an example of how to boot the DGF-4C modules using the C Library functions.

Table 7: An Example Code Illustrating How to Boot DGF-4C Modules.

```
// download module global names
C_Dgf4c_Hand_Down_Names(Module_Global_Names, "MODULE_GLOBAL_NAMES");
// download global data names
C_Dgf4c_Hand_Down_Names(Global_Data_Names, "GLOBAL_DATA_NAMES");
// download user variable names
C_Dgf4c_Hand_Down_Names(User_Var_Names, "USER_VAR_NAMES");
// initialize module global values
C_Dgf4c_User_Par_IO(Module_Global_Values, "MODULE_GLOBAL_VALUES", 0);
C_Dgf4c_Hand_Down_Names(All_Files, "ALL_FILES");  // download file names
C_Dgf4c_Boot_System(0x1F);  // boot DGF-4C modules
C_Dgf4c_Set_Current_ModChan(1,0);  // set current module and channel number
```

3.2 Setting DSP variables

The host computer communicates with the DSP by setting and reading a set of variables called DSP I/O variables. These variables, totally 416 unsigned 16-bit integers, sit in the first 416 words of the data memory. The first 256 words, which stores input variables, are both readable and writeable, while the remaining 160 words, which stores pointers to various data buffers and run summary data, are only readable. The exact location of any particular variable in the DSP code will vary from one code version to another. To facilitate writing robust user code, we provide a reference table of variable names and addresses with each DSP code version. Included with your software distribution is a file called DGFcodeE.var. It contains a two-column list of variable names and their respective addresses. Thus you can write your code such that it addresses the DSP variables by name, rather than by fixed location.

It should come as no surprise that many of the DSP variables have meaningful values and ranges depending on the values of other variables. A complete description of all interdependencies can be found in Section 4. All of these interdependencies have been taken care of by the DGF-4C C Library. So instead of directly setting DSP variables, users only need to set the values of the global variables defined in Table 4. The C Library will then convert these values into corresponding DSP variable values and download them into the DSP data memory. On the other hand, if users want to read out the data memory, the C Library will first convert these DSP values into the global variable values. The following code is an example of setting DSP variables through the C Library.

Table 8: An Example Code to Illustrating How to Set DSP Variables.

```
// set global data variable MODULE_CSRA to 0x2400
Global_Data_Values[Find_Xact_Global_DATA_Match("MODULE_CSRA")]=0x2400;
// download MODULE_CSRA to DSP
C_Dgf4c_User_Par_IO(Global_Data_Values, " GLOBAL_DATA_NAMES", 0);
// set user variable ENERGY_RISETIME to 6.0 μs
User_Var_Values[Find_Xact_User_Match ("ENERGY_RISETIME")]=6.0;
// download ENERGY_RISETIME to DSP
C_Dgf4c_User_Par_IO(User_Var_Values, "ENERGY_RISETIME", 0);
```

3.3 Access spectrum memory or list mode data

3.3.1 Access spectrum memory

The MCA spectrum memory is fixed to 32k words (24 bit) per channel, residing in the external memory. The memory is organized into 8 pages of 4k words. To read out the spectra to the host, each page has first to be transferred from the external memory to the I/O data buffer in the DSP memory, and then read out by a DMA transfer. The Spectrum memory is accessible after a MCA run, or a list mode run if histogramming energy is requested. The following code in Table 9 is an example of how to start a MCA run and read out the MCA spectrum after the run is finished.

Table 9: Accessing the spectrum memory.

```
// start a MCA run; dummy is an unsigned 32-bit integer array of any size
C_Dgf4c_Acquire_Data(0x1301, dummy, " ");
// wait until run has ended
while(C_Dgf4c_Acquire_Data(0x4000, dummy, " ")) {; }
// save MCA spectrum to a file
C_Dgf4c_Acquire_Data(0x3301, dummy, file_name);
// Read out the MCA spectrum and put it to array User_data
C_Dgf4c_Acquire_Data(0x7000, User_data, " ");
```

3.3.2 Access list mode data

The list mode data in the linear output data buffer can be written in a number of formats. User code should access the three variables BUFHEADLEN, EVENTHEADLEN, and CHANHEADLEN in the configuration file of a particular run to navigate through the data set.

There are two data buffers to choose from: the linear 8192 words long I/O buffer, and a circular level-1 data buffer with room for 2048 words. The buffer for the raw data is chosen automatically according to the run type. Note that because the level-1 buffer is shorter, the total amount of trace data that it can accept is limited to about 50µs.

In a multi-parametric setup, you can do any number of runs in a row. The first run would be started as a NEW run. This clears all histograms in memory. Subsequent runs would complete when the linear data buffer is full. Once it has been read out, you can RESUME running. This keeps the histogram memory intact and you can accumulate spectra over many runs. The example code shown in Table 10 illustrates this.

Table 10: Command sequence for multiple runs in a row.

```
C_Dgf4c_Acquire_Data(0x1100, dummy, " ");      // start a new list mode run
k = 1;     // initialize counter
do{
    while(C_Dgf4c_Acquire_Data(0x4000, dummy, " ")) {; }   // wait until run has ended
    C_Dgf4c_Acquire_Data(0x6000, dummy, file_name_1);     // stop run and save list mode run data
    k = k + 1;
    if(k > Nruns)
        break;
    C_Dgf4c_Acquire_Data(0x2100, dummy, " "));     // issue ResumeRun command
}while(1)
C_Dgf4c_Acquire_Data(0x3000, " ", file_name_2);       // store energy histogram
```

In response to a NEW run, the DSP will re-evaluate all parameter settings which may affect its data taking procedures, and the DSP will clear the I/O buffer, the level-1 buffer and the spectrum

memory, before beginning the actual data acquisition. In the revision-D modules this may take up to 4ms.

When receiving a RESUME run request, the DSP assumes that all DGF settings remain unchanged from the previous run. It will also refrain from erasing the spectrum memory, and will reset the global read and write pointers, which is the equivalent of flushing the data buffers, without physically zeroing out their contents. A RESUME run request is followed by actual data acquisition within about 10μs.

The DSP code can also create a compressed output data format. If the raw data is directed to the level-1 buffer, the DSP has access to pulse shapes. However, only energies and time stamps of the final data are written to the output buffer; all other event data are suppressed. This allows a more compact output format, which helps to reduce the necessary readout bandwidth.

To access the list mode data after the run is finished, the DGF-4C C Library provides several utility routines to parse the list mode data saved in the output file and read out the waveform or energy of each individual trace. The code in Table 11 shows how to read waveforms from a list mode file.

Table 11: An Example Code Showing How to Access List Mode Data.

```
C_Dgf4c_Acquire_Data(0x1100, dummy, " ");      // start a new list mode run
while(C_Dgf4c_Acquire_Data(0x4000, dummy, " ")) {; }   // wait until run has ended
C_Dgf4c_Acquire_Data(0x6000, dummy, file_name_1);     // store list mode data in a file
C_Dgf4c_Acquire_Data(0x3000, dummy, file_name_2);     // store energy histogram in a file
C_Dgf4c_Acquire_Data(0x5000, listmodewave, file_name_1);         // parse list mode file
totaltraces = 0;
for(i=0; i<2; i++)
    totaltraces += listmodewave[i+24];      // sum the total number of traces for the two modules
traceposlen = (long)malloc(totaltraces*3*4);     // allocate memory to hold position and length infomation
C_Dgf4c_Acquire_Data(0x5001, traceposlen, file_name_1);         // locate traces
Trace0 = (unsigned short)malloc(traceposlen[1]+2);        // allocate memory to hold the first trace
Trace0[0] = traceposlen[0];    // position of the first trace
Trace0[1] = traceposlen[1];    // length of the first trace
C_Dgf4c_Acquire_Data(0x5002, trace0, file_name_1);       // read out the first trace and put it into trace0
```

4 User Accessible Variables

User parameters are stored in the data memory space of the on-board DSP. The organization is that of a linear memory with 16-bit words. Subsequent memory locations are indicated by increasing addresses. The data memory space, as seen by the host computer, starts at 0x4000.

There are two sets of user-accessible parameters. 256 words in data memory are used to store input parameters. These can and must be set properly by the user application. A second set of 160 words is used for results furnished by the DGF-4C module. These should not be overwritten.

As of this writing the start address for the input parameter block is InParAddr=0x4000 and for the output parameter block it is OutParAddr=0x4100, i.e. the two blocks are contiguous in memory space. We provide an ASCII file named DGFcodeE.var which contains in a 2-column format the offset and name of every user accessible variable. We suggest that user code use this information to create a name→address lookup table, rather than relying on the parameters retaining their address offsets with respect to the start address.

The input parameter block is partitioned into 5 subunits. The first contains 64 data that pertain to the DGF-4C as a whole. It is followed by four blocks of 48 words, which describe the settings of the four channels.

Below we describe the module and channel parameters in turn. Where appropriate, we show how a variable can be viewed using the DGF4C-Viewer.

4.1 Module parameters

MODNUM: Logical number of the module. This number will be written into the header of the list mode buffer to aid offline event reconstruction.

MODCSRA: Module Control and Status Register A. The viewer panel is the Module CSRA Edit Panel. This is a bit-oriented variable.

Bit 0: Write data to level-1 buffer. In this mode data are first being written into a 2048 word deep buffer. Results, but not traces, are written to the list mode buffer. This mode is useful when only the results from pulse shape analyses, but not the traces, are requested.

Note that since software revision 2.60, the RUNTASK variable determines which buffer is the intermediate buffer in a run.

Bit 1..9: Reserved.
Set to 0.

Bit 10,13: Bitmask for switchbus settings, for revision-D and revision-E DGF-4Cs only. These bits are stored in DSP memory, but have to be written to the ICSR register by the host computer to set the FET switches for proper trigger line termination. Refer to section 9.3 of the DGF User's Manual for application details.

To terminate the fast trigger bus line with 100 Ω set bit 10 of the ICSR. To terminate the event trigger (dsp trigger) bus line with 100 Ω set bit 13 of the ICSR.

Bit 11..12 Reserved

Bit 14..15: Reserved.

MODCSRB: Module Control and Status Register B. This is a bit-oriented variable.

Bit 0: If set, user written DSP code is called.

Bit 1..15: Reserved.
Set to 0.

MODFORMAT: List mode data format descriptor. Currently it is not in use.

SUMDAC: The value of this variable controls the SUMDAC digital to analog converter. It can be used to set a trigger threshold on either the multiplicity or the sum analog signal from any combination of the four input channels. See the user's manual for the required jumper settings.

The threshold in volt can be calculated as follows:
Threshold = ((32768 - SUMDAC) / 32768)*3.0V

RUNTASK: This variable tells the DGF what kind of run to start in response to a run start request. Nine run tasks are currently supported.

RunTask	Mode	Trace Capture	CHANHEADLEN
0	**Slow control run**	**N/A**	**N/A**
256 (0x100)	**Standard list mode**	**Yes**	**9**
257 (0x101)	**Compressed list mode**	**Yes**	**9**
258 (0x102)	**Compressed list mode**	**Yes**	**4**
259 (0x103)	**Compressed list mode**	**Yes**	**2**
512 (0x200)	**Standard fast list mode**	**No**	**9**
513 (0x201)	**Compressed fast list mode**	**No**	**9**
514 (0x202)	**Compressed fast list mode**	**No**	**4**
515 (0x203)	**Compressed fast list mode**	**No**	**2**
769 (0x301)	**MCA mode**	**No**	**N/A**

RunTask 0 is used to request slow control tasks. These include programming the trigger/filter FPGAs, setting the DACs in the system, transfers to/from the external memory, and calibration tasks.

RunTask 256 (0x100) requests a standard list mode run. In this run type all bells and whistles are available. The scope of event processing includes computing energies to 16-bit accuracy, and performing pulse shape analyses for improved energy resolution and better time of arrival measurements. Nine words of results, including time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis, GSLTtimeA, GSLTtimeB, GSLTtimeC, etc. are written into the I/O buffer for each channel. Level-1 buffer is not used in this RunTask.

RunTask 257 (0x101) requests a compressed list mode run. Both Level-1 buffer and I/O buffer are used in this RunTask, but no traces are written into the I/O buffer. Nine words of results, including time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis, GSLTtimeA, GSLTtimeB, GSLTtimeC, etc. are written into the I/O buffer for each channel.

RunTask 258 (0x102) requests a compressed list mode run. The only difference between RunTask 258 and 257 is that in RunTask 258, only four words of results (time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis) are written into the I/O buffer for each channel.

RunTask 259 (0x103) requests a compressed list mode run. The only difference between RunTask 259 and 257 is that in RunTask 259, only two words of results (time of arrival and energy) are written into the I/O buffer for each channel.

RunTask 512 (0x200) employs the same internal data format as RunTask 256, but omits buffer-full checks and trace capture. The run is stopped when the required number of events (MaxEvents) has been acquired. This run type uses the shortest possible interrupt routine for raw data gathering. Hence, it allows for the shortest time between two logged events. For best results the channel variables PAFLength and TriggerDelay should be set to 1 for all channels involved. Level-1 buffer is not used in this run type. Nine words of results, including time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis, GSLTtimeA, GSLTtimeB, GSLTtimeC, etc. are written into the I/O buffer for each channel.

RunTask 513 (0x201) requests a compressed fast list mode run without trace capture. Both Level-1 buffer and I/O buffer are used in this RunTask. Nine words of results, including time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis, GSLTtimeA, GSLTtimeB, GSLTtimeC, etc. are written into the I/O buffer for each channel.

RunTask 514 (0x202) requests a compressed fast list mode run. The only difference between RunTask 514 and 513 is that in RunTask 514, only four words of results

(time of arrival, energy, XIA pulse shape analysis, user pulse shape analysis) are written into the I/O buffer for each channel.

RunTask 515 (0x203) requests a compressed fast list mode run. The only difference between RunTask 515 and 513 is that in RunTask 515, only two words of results (time of arrival and energy) are written into the I/O buffer for each channel.

RunTask 769 (0x301) requests a MCA run. The raw data stream is always sent to the level-1 buffer, independent of MODCSRA. The data-gathering interrupt routine fills that buffer with raw data, while the event processing routine removes events after processing. If the interrupt routine finds the level-1 buffer to be full, it will ignore events until there is again room in the buffer. The run will not abort due to buffer-full condition. This run type does not write data to the I/O buffer. The module variable MAXEVENTS should be set to zero, to avoid early run termination due to a MAXEVENTS-exceeded condition.

The RunTask can be chosen as the run type in the Run tab of the DGF-4C Viewer.

CONTROLTASK: Use this variable to select a control task. Consult the control tasks section of this manual for detailed information. The control task will be launched when you issue a run start command with RUNTASK=0.

MAXEVENTS: The module ends its run when this number of events has been acquired. In DGF-4C Viewer, MAXEVENTS is automatically calculated when a run mode is chosen from the run type pulldown menu. The calculation is based on the trace lengths set by the user. Set MaxEvents=0 if you want to switch off this feature, e.g., when logging spectra (done automatically in an MCA mode run).

COINCPATTERN: When a DGF-4C is operated on its own, the user can request that certain coincidence/anticoincidence patterns are found for the event to be accepted. With four channels there are 16 different hit patterns, and each can be individually selected or marked for rejection by setting the appropriate bit in the COINCPATTERN mask.

Consider the 4-bit hit pattern 1010. The two 1's indicate that channel 3 (MSB) and channel 1 have reported a hit. Channels 2 and 0 did not. The 4-bit word reads as 10(decimal). If this hit pattern qualifies as an acceptable event, set bit 10 in the COINCPATTERN to 1. The 16 bit in COINCPATTERN cover all combinations. Setting COINCPATTERN to 0xFFFF causes the DGF to accept any hit pattern as valid.

In the DGF-4C Viewer this variable can be set in the Coincidence Pattern Edit Panel reachable through the Settings tab by clicking on Edit next to the Coinc. Pattern entry.

COINCWAIT: Duration of the coincidence time window in 25ns clock ticks. The actual coincidence window is 50ns wider than the value determined by COINCWAIT. For this feature to work, bit no. 1 of the ChannelCSRA of the involved channels should be cleared. This ensures that the DSP can at the end of the coincidence window suppress further hits reporting by late channels.

In the DGF-4C Viewer this bit is set or cleared in line 1 of the Channel CSRA Edit Panel. The line has the title "Measure individual live time". Make sure it is unchecked, so the DSP globally controls FPGA triggering and live time measurements.

When acquiring long waveforms it may be necessary to delay DSP data reading to ensure that the FIFOs will contain valid data. Secondly, when using 6-bit decimation in the FPGA, the minimum value for COINCWAIT is larger than 1 in all circumstances. Use the following formula to determine COINCWAIT:

COINCWAIT = max(1,ceil(1.5*2^Decimation-70))
COINCWAIT = max(COINCWAIT, PAFLENGTH-TRIGGERDELAY-70)

Choose COINCWAIT big enough such that the requirements of all channels in the module are met.

SYNCHWAIT: Controls run start behavior. When set to 0 the module simply starts or resumes a run in response to the corresponding request. When set to 1, closing the Busy--Synch loop is required. For a single module this is accomplished by connecting the Busy output to the Synch input via a Lemo cable. If two or more modules are in the system and are to run synchronously, a more complex wiring scheme is needed. All Busy outputs must lead to the inputs of a multi-input OR. The result from the OR operation must be fed back to the Synch inputs. This kind of set up in connection with SyncWait=1 will ensure that the last module ready to actually begin data taking will start the run in all modules. And the first module to end the run will stop the run in all modules. This way it never happens that a multi-DGF system is only partially active.

INSYNCH: InSynch is an input/output variable. It is used in multi-DGF systems in which the modules are driven by a common clock. When InSynch is 1, the module assumes it is in synch with the other modules and no particular action is taken at run start. If this variable is 0, then all system timers are cleared at the beginning of the next data acquisition run (RunTask>0). Using the Busy-Synch loop as described above, the timers are reset when the entire system actually starts the run. After run start, InSynch is automatically set to 1. Clock resetting can occur only if the Busy--Synch loop is closed.

HOSTIO: A 4 word data block that is used to specify command options. Currently it is only used to

- specify the channel in an external memory transfer;

- specify the channel when reading untriggered traces;
- specify the channel for baseline measurements.

XdatLength: Length of a data block to be downloaded from the host. Use XdatLength=0 as the default value for normal operation.

USERIN: A block of 16 input variables used by user-written DSP code.

U00: Many unused, but reserved, data blocks have names of the structure Unn. Those unused data blocks which reside in the block of input parameters for each channel are called UNUSEDA and UNUSEDB.

UNUSEDA: Only used in Controltask 4 for reading untriggered traces. UNUSEDA stores the weight in the geometric-weight averaging scheme to remove higher frequency signal and noise components. The value is calculated as follows:
For a given dt (in μs), calculate the integer intdt = dt/0.025
Then, if intdt>=11, XWAIT=4*floor((intdt-3)/4)+3
Finally, UNUSEDA = floor(65536/((intdt-3)/4))
If intdt<11, UNUSEDA is ignored.

4.2 Channel variables

All channel-0 variables end with "0", channel-1 variables end with "1", etc. In the following explanations the numerical suffix has been removed. Thus, e.g., CHANCSRA0 becomes CHANCSRA, etc.

CHANCSRA: The control and status register bits switch on/off various aspects of the DGF-4C operation, see the CCSRAeditPanel reachable through the InstrumentPanel of the DGF-4C Viewer. In general, setting the bit activates the option in question.

Bit 0: Respond to group triggers only.
Set this bit if you want to control the waveform acquisition for non-triggering channels by a triggering master channel. For this option to work properly choose one channel as the master and have its Trigger_Enable bit set. All dependent channels should have their Trigger_Enable bit cleared. Set bit 0 in all slave channels. You should also set it the master channel to ensure equal time of arrivals for the fast trigger signal, which is used to halt the FIFOs.

Bit 1: Measure individual live time.
Keep this bit cleared when operating with master and slave channels, or when making coincidence measurements using single modules. Set this bit when measuring independent spectra, i.e., when list mode data are not required.

Bit 2: Good channel.
Only channels marked as good will contribute to spectra and list mode data.

Bit 3: Read always
Channels marked as such will contribute to list mode data, even if they did not report a hit. This is most useful when acquiring induced signal waveforms on spectator electrodes, i.e., electrodes that did not collect any net charge, but only saw a transient induced signal.

Bit 4: Enable trigger.
Set this bit for channels that are supposed to contribute to an event trigger.

Bit 5: Trigger positive.
Set this bit to trigger on a positive slope; clear it for triggering on a negative slope. The trigger/filter FPGA can only handle positive signals. The DGF handles negative signals by inverting them immediately after entering the FPGA.

Bit 6: GFLT.
Set this bit if you want to validate or veto events using the front panel GFLT LEMO input. When the bit is cleared, the GFLT input is ignored. When set, the event is accepted only if validated. To be validated, the GFLT input must be a logic 1 no later than an energy filter rise time after the signal arrival, and must remain at logic 1 level until a rise time + flat top after signal arrival.

Bit 7: Histogram energies.
Set this bit to histogram energies from this channel in the on-board MCA memory.

Bit 8: Reserved.
Set to 0.

Bit 9: Reserved.

Bit 10: Compute constant fraction timing.
This pulse shape analysis computes the time of arrival for the signal from the recorded waveform. The result is stated in units of $1/256^{th}$ of a sampling period (25ns). Time zero is the start of the waveform.

Bit 11: Enable contribution to multiplicity.
Any of the four channels can contribute to the multiplicity output at the front panel. With this bit one can switch this contribution on or off.

Bit 12..15: Reserved.

CHANCSRB: Control and status register B.

Bit 0: If set, call user written DSP code.

Bit 1: If set, all words in the channel header except Ndata, trigtime and energy will be overwritten with the contents of URETVAL. Depending on the run type, this allows for 6, 2 or 0 user return values in the channel header.

Bit2..15: are reserved. Set to 0.

The following two data words are used to set the on-board DACs for this channel. Once a new variable has been written to DSP memory the DACs have to be reprogrammed by starting a run with RunTask=0 and ControlTask=0.

GAINDAC: This DAC is used to program the variable gain amplifier. The GainDAC value corresponds to a gain according to the following formula.

Gain [V/V] = 0.1639 * 10^((65535 - GAINDAC) / 32768)

TRACKDAC: This DAC determines the DC-offset voltage. The offset can be calculated using the following formula:

Offset [V] = 3.0 * ((32768 - TRACKDAC) / 32768)

U02: Begin of a reserved data block.

The following block of data contains trigger/filter FPGA data. Once a new variable has been written to DSP memory it has to be activated by starting a run with RunTask 0 and ControlTask 5.

SLOWLENGTH: The rise time of the energy filter depends on SlowLength:

RiseTime = SlowLength * 2^Decimation * 25ns

SLOWGAP: The flat top of the energy filter depends on SlowGap: FlatTop = SlowGap * 2^Decimation * 25ns.

There is a constraint concerning the sum value of the two parameters:

SlowLength + SlowGap < 32

FASTLENGTH: The rise time of the trigger filter depends on FastLength: RiseTime = FastLength * 25ns.

Note the constraint: FastLength < 32

FASTGAP: The flat top of the trigger filter depends on FastGap: FlatTop = FastGap * 25ns.

There is a constraint concerning the sum value of the two parameters: FastLength +

FastGap < 32

FASTADCTHR: This value is used by the DGF-4C Viewer to store the fast trigger threshold in ADC units. The DGF-4C module does not use this value.

For **revision-D** and **revision-E** modules one LSB of this variable corresponds to 4 LSB of the reported waveform data, which are 14-bit numbers.

FASTTHRESH: This is the trigger threshold used by the trigger/filter FPGA. The value relates to a trigger threshold in ADC units through the formula:

FASTTHRESH = FastADCthr * FASTLENGTH

MINWIDTH: This value aids the pile up inspector. MinWidth is the minimum duration, in sample clock ticks (25ns), which the output from the fast filter must spend over threshold. Pulses shorter than that will be rejected as noise spikes. The recommended setting is MinWidth = FastLength + FastGap

MAXWIDTH: This value aids the pile up inspector. MaxWidth is the maximum duration, in sample clock ticks (25ns), which the output from the fast filter may spend over threshold. Pulses longer than that will be rejected as piled up. The recommended setting is MaxWidth = FastLength + FastGap + SignalRiseTime/25ns.

Note the constraint MaxWidth < 256.

Setting Maxwidth=0 switches this part of the pile up inspector off. Indeed it is recommended to begin with MaxWidth=0. Once the other parameters have been optimized, one can use the MaxWidth cut to improve the pile up rejection at high count rates. Maxwidth should be tuned by observing the main energy peak in the spectrum for fixed time intervals. Once the MaxWidth cut is too tight there will be a loss of efficiency in the main peak. Setting MaxWidth to such a value that the efficiency loss in the main peak is acceptable will give the best overall performance in terms of efficiency and pile up rejection.

PEAKSAMPLE: This variable determines at what time the value from the energy filter will be sampled. Note that the following formulae depend on the decimation:

0-bit decimation: PeakSample = max(0, SlowLength + Slow Gap – 7)
1-bit decimation: PeakSample = max(2, SlowLength + Slow Gap – 4)
2-bit decimation: PeakSample = SlowLength + Slow Gap – 2
3-bit and higher decimation: PeakSample = SlowLength + Slow Gap – 1

If the sampling point is chosen poorly, the resulting spectrum will show energy resolutions of 10% and wider rather than the expected fraction of a percent. For some parameter combinations PeakSample needs to be varied by one or two units in either direction, due to the pipelined architecture of the trigger/filter FPGA.

PEAKSEP: This value governs the minimum time separation between two pulses. Two pulses that arrive within a time span shorter than determined by PeakSep will be rejected as piled up.

The recommended value is: PeakSep = PeakSample+5
If PeakSep>33, PeakSep=PeakSample+1

Note the constraint: 0 < PeakSep - PeakSample < 7.

PAFLENGTH: A FIFO control variable that needs to be written into the trigger/ filter FPGA. Using the programmable almost-full register we can time the waveform capturing thus that by the time the DSP is triggered at the end of the pile up inspection period the data of interest have percolated through to the begin of the FIFO and are available for read out without delay.

The acquired waveform will start rising from the baseline at a time delay after the beginning of the trace. This delay is a quantity that the user will want to set. In the DGF-4C Viewer it is called Trace Delay (measured in microseconds) and is available through the Settings tab.

The recommended setting for PafLength is:

PafLength = TriggerDelay + UserDelay/0.025 + 8

Note the constraint: PafLength < 4092.

Note that PAFLength should be adjusted only in multiples of 4, as the hardware ignores the lower two bits of this value.

TRIGGERDELAY: This is a partner variable to PafLength. For *all* decimations,

TriggerDelay = (PeakSample+6)*2^(Decimation)

Note that TriggerDelay should be adjusted only in multiples of 4, as the hardware ignores the lower two bits of this value. For MCA runs without taking traces, (trace length=0), TriggerDelay should be 1.

RESETDELAY: This variable controls the restarting of the FIFO after it was halted to read the waveform. When triggers are distributed across channels and modules, a halted FIFO is automatically restarted if the trigger/filter FPGA does not receive the distributed event trigger within RESETDELAY 25ns clock ticks after the internal event trigger. The default value written by the DGF module should not be changed by the user.

FTPWIDTH: The fast trigger pulse, which is sent to the multiplicity output has a programmable width, set through FTPWidth. The pulse width is given in sampling clock periods of 25ns.

Note the constraint: FTPwidth < 256.

This completes the list of values that control the trigger/filter FPGAs.

The following input parameters are used by the DSP program. They become active as soon as the first data taking run has been started. Only then will the output parameters reflect the changes made to the set of input parameters.

TRACELENGTH: This tells the DSP how many words of trace data to read. The action taken depends on FIFOlength, which is 4096 for all Rev-D and Rev-E modules. If TraceLength < FIFOlength, the DSP will read from the FIFO. In that case individual samples are 25ns apart. If FIFOlength <= TraceLength, the DSP will read from an FPGA register which mirrors the current ADC output. In that case individual readings are 75ns apart. In addition only post trigger data will be available because the DSP is then reading data in real time rather than data stored in a FIFO. TraceLengths greater than FIFOlength are useful only when reading out only one channel. In this case it allows acquiring a long trace to measure the exponential decay time of a preamplifier.

XWAIT: Extra wait states. This parameter controls how many extra clock cycles the DSP waits when reading waveform data in real time rather than out of a FIFO memory. This occurs when acquiring data in list mode and asking for trace lengths longer than FIFOlength. The time between recorded samples is

$\Delta T = (3 + XWAIT)*25ns.$

XWAIT is used differently when acquiring untriggered traces in a control run with ControlTask=4. In this case, the time between recorded samples is

$\Delta T =$	3*25ns	if XWAIT <= 3;
	4*25ns	if XWAIT = 4;
	5*25ns	if XWAIT = 5;
	(3 + XWAIT)*25ns	if 5 < XWAIT <= 11;
	(3 + XWAIT)*25ns	if XWAIT > 11 (XWAIT has to be multiple of 4)

The following variables affect internal MCA histogramming of the DGF-4C module.

ENERGYLOW: Start energy histogram at ENERGYLOW

LOG2EBIN: This variable controls the binning of the histogram. Energy values are calculated to 16 bits precision. The LSB corresponds to $1/16^{th}$ of a 12-bit ADC unit (or $1/4^{th}$ of a

14-bit ADC for revision-E modules). The DGFs, however, do not have enough histogram memory available to record 64k spectra, nor would this always be desirable. The user is therefore free to choose a lower cutoff for the spectrum (EnergyLow) and control the binning. Observe the following formula to find to which MCA bin a value of Energy will contribute:

MCAbin = (Energy-EnergyLow) * 2^Log2Ebin

As can be seen, Log2Ebin should be a negative number to achieve the correct behaviour. At run start the DSP program ensures that Log2Ebin is indeed negative by replacing the stored value by -abs(Log2Ebin).

The histogramming routine of the DSP takes care of spectrum overflows and underflows.

CFDTHR: This sets the threshold of the software constant fraction discriminator. The threshold fraction (f) is encoded as Round(f*65536), with 0<f<1.

PSAOFFSET:
PSALENGTH: When recording traces and requiring any pulse shape analysis by the DSP, these two parameters govern the range over which the analysis will be applied. The analysis begins at a point PSAOFFSET sampling clock ticks into the trace, and is applied over a piece of the trace with a total length of PSALENGTH clock ticks.

INTEGRATOR: unused, reserved

BLCUT: This variable sets the cutoff value for baselines in baseline measurements. If BLCUT is not set to zero, the DSP checks continuously each baseline value to see if it is outside of the limit set by BLCUT. If the baseline value is within the limit, it will be used to calculate the average baseline value. Otherwise, it will be discarded. Set BLCUT to zero to not check baselines, therefore reduce processing time.

ControlTask 6 can be used to measure baselines. Host computer can then histogram these baseline values and determine the appropriate value for BLCUT for each channel according to the standard deviation SIGMA for the averaged baseline value. BLCUT could be set to be three times SIGMA.

CFDREG: Reserved for FPGA-based constant fraction discriminator.

LOG2BWEIGHT: The DGF measures baselines continuously and effectively extracts DC-offsets from these measurements. The DC-offset value is needed to apply a correction to the computed energies. To reduce the noise contribution from this correction baseline samples are averaged in a geometric weight scheme. The averaging depends on Log2Bweight:

DC_avg = DC + (DC_avg-DC) * 2^LOG2BWEIGHT

DC is the latest measurement and DC_avg is the average that is continuously being updated. At the beginning, and at the resuming, of a run, DC_avg is seeded with the first available DC measurement.

As before, the DSP ensures that LOG2BWEIGHT will be negative. The noise contribution from the DC-offset correction falls with increased averaging. The standard deviation of DC_avg falls in proportion to sqrt(2^LOG2BWEIGHT).

When using a BLCUT value from a noise measurement (cf control task 19) the DGF will internally adjust the effective Log2Bweght for best energy resolution, up to the maximum value given by LOG2BWEIGHT. Hence, the Log2Bweight setting should be chosen at low count rates (dead time < 10%). Best energy resolutions are typically obtained at values of -3 to -4, and this parameter does not need to be adjusted afterwards.

U04: Begin of an unused data block.

PREAMPTAUA: High word of the preamplifier exponential decay time.
PREAMPTAUB: Low word of the above.
The two variables are used to store the preamplifier decay time. The time τ is measured in μs. The two words are computed as follows.

PREAMPTAUA = floor(τ)
PREAMPTAUB = 65536 * (τ - PreampTauA)
To recover τ use:
τ = PREAMPTAUA + PREAMPTAUB / 65536

This ends the block of channel input data. Note that there are four equivalent blocks of input channel data, one for each DGF-4C input channel.

We now show the output variables, again beginning with module variables and continuing afterwards with the channel variables. The output data block begins at the address 0x4100. Note, however, that this address could change. The output data block comprises of 160 words; 1 block of 32 is reserved for module data; 4 blocks of 32 words each hold channel data.

DECIMATION: The DSP reads this value from the trigger/filter FPGA. It is a characteristic of the configuration that was downloaded. You will find that the available configuration files support decimations of 0, 1, 2, 3, 4, 5 and 6.

REALTIMEA:
REALTIMEB:
REALTIMEC: The 48-bit real time clock. A,B,C are the high, middle and low word, respectively. The clock is zeroed on power up, and in response to a synch interrupt when InSynch was set to 0 prior to the run start. This requires the Busy--Synch loop

to be closed; see the discussion above.

RealTime =(RealTimeA * 65536^2 + RealTimeB * 65536 + RealTimeC) * 25ns

RUNTIMEA:
RUNTIMEB:
RUNTIMEC: The 48-bit run time clock. A,B,C words are as for the RealTime clock. This time counter is active only while a data acquisition run is in progress. Comparing the run time with the real time allows judging the overhead due to data readout. Compute the run time using the following formula:

RunTime =(RunTimeA * 65536^2 + RunTimeB * 65536 + RunTimeC) * 25ns

GSLTTIMEA:
GSLTTIMEB:
GSLTTIMEC: Signal arrival time of a logic 0→1 transition at the GSLT front panel LEMO input. The time latched is the real time at that moment.

NUMEVENTSA:
NUMEVENTSB: Number of valid events serviced by the DSP.
Again the high word carries the suffix A and the low word the suffix B.

DSPERROR: This variable reports error conditions:
= 0 (NOERROR), no error
= 1 (RUNTYPEERROR), unsupported RunType
= 2 (RAMPDACERROR), Baseline measurement failed

SYNCHDONE: This variable can be set to 1 to force the DSP out of an infinite loop caused by a malfunctioning Busy-Synch loop, when a run start request was issued with SYNCHWAIT=1.

BUFHEADLEN: At the beginning of each run the DSP writes a buffer header to the list mode data buffer. BufHeadLen is the length of that header. Currently, BUFHEADLEN is 6, but this value should not be hardcoded, it should be read from the DSP to ensure upgrade compatibility.

EVENTHEADLEN: For each event in the list mode buffer, or the level-1 buffer, there is an event header containing time and hit pattern information. EventHeadLen is the length of that header. Currently, EVENTHEADLEN is 3, but this value should not be hardcoded, it should be read from the DSP to ensure upgrade compatibility.

CHANHEADLEN: For each channel that has been read, there is a channel header containing energy and auxiliary information. ChanHeadLen is the length of this header. CHANHEADLEN varies between 2 and 9 words depending on the run type (see RUNTASK).

The event and channel header lengths plus the requested trace lengths determine the maximum logically possible event size. The maximum event size is the sum of EventHeadLen and the ChannelHeadLengths plus the TraceLengths for all channels marked as good, i.e. which have bit 2 in the ChanCSRA set. Example: With all four channels marked as good and required trace lengths of 1000 (i.e. 25μs) the maximum event size will be

MaxEventSize = EventHeadLen + 4*(ChanHeadLen + 1000)
= 4039

In the last line typical values for EventHeadLen (3) and ChanHeadLen (9) were substituted. BufHeadLen equals 6. Thus there is room for at least 2 events in the list mode data buffer, which is 8192 words long. But there is not enough room in the level-1 buffer, which contains only 2048 words.

Below follow the addresses and lengths of a number of data buffers used by the DSP program. The addresses are generated by the assembler/linker when creating the executable. On power up the DSP code makes these values accessible to the user. Note that the addresses will typically change with every new compilation. Therefore your code should never assume to find any given buffer at a fixed address.

Note that addresses in the DSP data memory fall into the range from 0x4000 to 0x7FFF. The word length in data memory is 16 bit. If an address falls in the range from 0 to 0x3FFF, it points to a location in program memory. Here the word lengths are 24 bits.

USEROUT: 16 words of user output data, which may be used by user written DSP code.

AOUTBUFFER: Address of the list mode data buffer.
LOUTBUFFER: Number of words in the list mode buffer.

AECORR: unused, reserved
LECORR: unused, reserved.

Formerly address and length of an array containing coefficients for energy calculations. Now these coefficients are calculated in the DSP code from the decay time.

ATCORR: unused, reserved
LTCORR: unused, reserved

Formerly address and length of an array containing coefficients for normalization and time of arrival corrections. Now these coefficients are calculated in the DSP code from the decay time.

HARDWAREID: ID of the hardware version
HARDVARIANT: Variant of the hardware
FIFOLENGTH: Length of the onboard FIFOs, measured in storage locations. Rev-D and Rev-E DGF-4Cs have FIFOs with 4096 locations.

FIPPIID: ID of the FiPPI FPGA configuration
FIPPIVARIANT: Variant of the FiPPI FPGA configuration

INTRFCID: ID of the CAMAC interface FPGA configuration
INTRFCVARIANT: Variant of the CAMAC interface FPGA configuration

DSPRELEASE: DSP software release number
DSPBUILD: DSP software build number

The following channel variables contain run statistics. Again the variable names carry the channel number as a suffix. For example the LIVETIME words for channel 2 are LIVETIMEA2, LIVETIMEB2, LIVETIMEC2. Channel numbers run from 0 to 3.

LIVETIMEA:
LIVETIMEB:
LIVETIMEC: Total live time as measured by the trigger/filter FPGA of that channel. It excludes times during which the FPGA was prevented from sending triggers due to ongoing DSP data reads, or when the run was stopped. Convert the three LiveTime words into a live time using the formula:

LiveTime = (LiveTimeA * 65536^2 + LiveTimeB * 65536 + LiveTimeC) * 0.400μs

FASTPEAKSA: The number of events detected by the fast filter is:
FASTPEAKSB: NumEvents = FASTPEAKSA*65536 + FASTPEAKSB

ADCPERDACA: Gain variable.
ADCPERDACB: Both words currently unused, but reserved

4.3 ADC data

The revision-E DGF-4C modules employ 14-bit waveform digitizing ADCs, while revision-D DGF-4C modules employ 12-bit ADCs. All modules are operating at 40MSPS. Hence, the natural units are 25ns for a time step. Depending on which DGF-4C modules being used, the original waveform data are either 14-bit unsigned numbers ranging from 0 to 16383 or 12-bit unsigned numbers ranging from 0 to 4095. Derived quantities, however, are reported by the DGF to higher than 12-bit precision.

Energy values are all reported as unsigned 16-bit numbers, and a pulse step covering the full range of the ADC would be reported as having amplitude of 65535. That is, an LSB of an energy value

corresponds to $1/4^{th}$ of an original ADC unit for 14-bit ADCs or to $1/16^{th}$ of an original ADC unit for 12-bit ADC units.

Waveform data are reported as untriggered traces in the Oscilloscope of the DGF4C-Viewer (cf control task 4), or as triggered traces in the list mode trace display of the DGF4C-Viewer during regular data acquisition. Revision-D and E DGFs report the waveforms in the Oscilloscope as 14-bit numbers and the list mode trace as 16-bit numbers.

The trigger threshold set by the FASTTHRESH variable is always in units of the 12-bit ADC times the length of the trigger filter measured in 25ns ticks.

The Ramp Offset DAC control task (#3) always reports results in units of the 12-bit ADC.

The following example may illuminate this. Assume a DC-offset such that it measures 400 units on the 12-bit ADC. Assume a step pulse with a pulse height of 1000 units as measured by the 12-bit ADC.

Oscilloscope shows baseline at 1600. The pulse step would show as a jump from 1600 to 5600 in the Oscilloscope display. Traces acquired in list mode data runs would show a jump from 6400 to 22400. The energy would be reported as 1000.

For a trigger filter length of 100ns (FastLength=4) and FASTTHRESH=100 the trigger threshold will be 25 units of the 12-bit ADC. This value is shown by the DGF4C-Viewer as the threshold value in the Settings tab.

5 Control Tasks

The DSP can execute a number of control tasks, which are necessary to control hardware blocks that are not directly accessible from the host computer. The most prominent tasks are those to set the DACs, program the trigger/filter FPGAs and read the histogram memory. The following is a list of control tasks that will be of interest to the programmer.

To start a control task, set RUNTASK=0 and choose a CONTROLTASK value from the list below. Then start a run by setting bit 0 in the control and status register (CSR).

Control tasks respond within a few hundred nanoseconds by setting the RUNACTIVE bit (#13) in the CSR. The host can poll the CSR and watch for the RUNACTIVE bit to be deasserted. All control tasks indicate task completion by clearing this bit.

Execution times vary considerably from task to task, ranging from under a microsecond to 10 seconds. Hence, polling the CSR is the most effective way to check for completion of a control task.

Control Task 0: **SetDACs**
Write the gaindac and trackdac values of all channels into the respective DACs. Also program the SumDAC. Reprogramming the DACs is required to make effective changes in the values of the variables GAINDAC{0…3}, TRACKDAC{0…3} and SUMDAC.

Control Task 1: **Connect inputs**
Close the input relay to connect the DGF electronics to the input connector.

Control Task 2: **Disconnect inputs**
Open the input relay to disconnect the DGF electronics from the input connector.

Control Task 3: **Ramp offset DAC**
This is used for calibrating the offset DAC. For each channel the offset DAC is incremented in 2048 equal-size steps. At each DAC setting the DC-offset is determined and written into the list mode buffer. At the end of the task the list mode buffer holds the following data. Its 8192 words are divided up equally amongst the four channels. Data for channel 0 occupy the lowest 2048 words, followed by data for channel 1, etc. The first entry for each channel's data block is for a DAC value of 0, the last entry is for a DAC value of 65504. In between entries the DAC value is incremented in steps of 32.

An examination of the results will reveal a linearly rising or falling response of the ADC to the DAC increments. The slope depends on the trigger polarity setting, i.e., bit 5 of the channel control and status register A

(ChanCSRA). For very low and very big DAC values the ADC will be driven out of range and an unpredictable, but constant response is seen. From the sloped parts a user program can find the DAC value that is necessary for a desired ADC offset. It is recommended, that for unipolar signals an ADC offset of 400 units is chosen. For bipolar signals, like the induced waveforms from a segmented detector, the ADC offset would be 2048 units, i.e., midway between 0 and 4095.

Note that for revision-D modules, ADC waveforms are reported as 14-bit numbers, ranging from 0 to 16383. Hence, the DC-offsets should be adjusted to produce readings of 1600 and 8192 counts, respectively, for unipolar and bipolar signals.

A user program would use the result from the calibration task to find, set and program the correct offset DAC values.

Since the offset measurement has to take the preamplifier offset into account, this measurement must be made with the preamplifier connected to the DGF input. The control task makes 16 measurements at each DAC step and uses the last computed DC-offset value to enter into the data buffer. Due to electronic noise, it may occasionally happen that none of the sixteen attempts at a base line measurement is successful, in which case a zero is returned. The user software must be able to cope with an occasional deviation from the expected straight line.

On exit, the task restores the offset DAC values to the values they had on entry.

ControlTask 4: **Untriggered Traces**

This task provides ADC values measured on all four channels and gives the user an idea of what the noise and the DC-levels in the system are. This function samples 8192 ADC words for the channel specified in HOSTIO. The XWAIT variable determines the time between successive ADC samples (samples are XWAIT * 25ns apart). In the DGF-4C Viewer XWAIT can be adjusted through the dT variable in the Oscilloscope panel. The results are written to the 8192 words long I/O buffer. Use this function to check if the offset adjustment was successful.

From the DGF-4C Viewer this function is available through the Oscilloscope Panel. Hit the Refresh button to start four consecutive runs with ControlTask 4 in the selected module, one for each channel.

ControlTask 5: **ProgramFiPPI**

Write all relevant data to the FiPPI control registers and download the ADC calibration data. The latter must be written into the I/O buffer (starting at the address given by the variable AOUTBUFFER) prior to calling this task. If

the ADC calibration data are not available, fill the first 1024 locations of the I/O buffer with zeros.

ControlTask 6: **Measure Baselines**

This routine is used to collect baseline values. Currently, DSP collects six words, B0L, B0H, B1L, B1H, time stamp, and ADC value, for each baseline. 1365 baselines are collected until the 8192-word I/O buffer is almost completely filled. The host computer can then read the I/O buffer and calculate the baseline according to the formula:

$B1 = (B1L+B1H*65536)/2^{(DECIMATION+9)}$
$B0 = (B0L+B0H*65536)/ 2^{(DECIMATION+9)}$
$TAU = PreampTauA+PreampTauB/65536$
$Baseline = B1-B0*\exp(-0.025*(SlowLength+SlowGap)*2^{DECIMATION}/TAU)$

Baseline values can then be statistically analyzed to determine the standard deviation associated with the averaged baseline value and to set the BLCUT. BLCUT should be about 3 times the standard deviation. Baseline values can also be plotted against time stamp or ADC value to explore the detector performance. BLCUT should be set to zero while running ControlTask 6.

ControlTask 9: **Read histogram memory, 1st page**

(for modules with serial numbers D1100 through D1199, E1200 and above) transfers the first page of external memory into the DSP's data memory. The target location in the DSP memory is the main I/O buffer, beginning at the address contained in AOUTBUFFER. The external memory is organized on a channel-by-channel basis; and HOSTIO specifies which channel to transfer.

This routine is used to read out spectra after a run. During a data run, each channel fills a 32k spectrum (24 bits deep) in the external memory. The external memory is organized into 8 pages of 4k words. In the control run, one (4k x 24 bits) page is written into the (8k x 16 bits) I/O buffer. After the control run is finished, the host has to read out the I/O buffer. ControlTask 9 always transfers the first page of a given channel, then increments a page counter. For subsequent transfers of the remaining 7 pages, use ControlTask 10.

ControlTask 10: **Read histogram memory, subsequent pages**

(for modules with serial numbers D1100 through D1199, E1200 and above) transfers a page of external memory into the DSP's main I/O buffer. An internal page counter specifies the page. It is set to zero (first page) in a ControlTaks 9, and is incremented by the DSP after each ControlTasks 9 or 10. To read out all 8 pages of a channels spectrum, first start a run with ControlTask 9, followed by seven runs with ControlTask 10, reading out the

I/O buffer after every run.

The 3 bytes of a histogram entry B0(LSB), B1 and B2 are mapped onto two 16-bit words W0(B1,B0) and W1(0x00,B2) in the I/O buffer. W0 occupies the lower memory address, and B0 and B2 form the least significant bytes of W0,W1.

ControlTask 11: **Write histogram memory, 1st page**
(for modules with serial numbers D1100 through D1199, E1200 and above) writes the DSP's main I/O buffer into the first page of external memory. The target channel is specified by HOSTIO.

ControlTask 12: **Write histogram memory, subsequent pages**
(for modules with serial numbers D1100 through D1199, E1200 and above) writes the DSP's main I/O buffer into a subsequent page of external memory. The page is specified by an internal page counter. The page counter is set to zero (first page) in a ControlTaks 11, and incremented after each ControlTask 11 or 12.

The byte organization is the same as described at the end of Control task 10.

ControlTask 19: **reserved**

ControlTask 20: **First ADC-Calibration,** (DGF-4C revision-D only)

ControlTask 21: **Subsequent ADC-Calibrations,** (DGF-4C revision-D only)
Contact factory for details.

6 Appendix A — User supplied DSP code

6.1 Introduction

It is possible for users to enhance the capabilities of the DGF-4C by adding their own DSP code. XIA provides an interface on the DSP level and has built support for this into the DGF-4C Viewer. The following sections describe the interfaces and support features.

6.2 The development environment

For the DSP code development, XIA uses and recommends version 5 or 6 of the assembler and linker distributed by Analog Devices. Both versions are in use at XIA and work fine.

It may be inconvenient, but is unavoidable to program the ADSP-2181 on board processor in assembler rather than in a higher level programming language like C. We found that code generated by the C-compiler is bloated and consequently runs very slow. As the main piece of the code could not be written in C at all, we did not burden our design by trying to be compatible with the C-compiler. Hence, using the C-compiler is currently not an option.

With the general software distribution we provide working executables and support files. To support user DSP programming we provide files containing pre-assembled forms of XIA's DSP code, together with a source code file that has templates for the user functions. The user templates have to be converted by the assembler and the whole project is brought together by the linker. XIA provides a link and a make file to assist the process.

In the DGF-4C Viewer we provide powerful diagnostic tools to aid code developing and a data interface to exchange data between the host and the user code. The DGF-4C Viewer can, at any time, examine the complete memory content of the DSP and call any variable from any code section by name. A particularly useful added feature is the capability to download data in native format into the DSP and pretend that they were just acquired. The event processing routine, which calls the user code, is then activated and processes the data. This in-situ code testing allows the most control in the debugging process and is more powerful than having to rely on real signal sources.

6.3 Interfacing user code to XIA's DSP code

When the DSP is booted it launches a general initialization routine to reach a known, and useful, state. As part of this process a routine called UserBegin is executed. It is used to communicate addresses and lengths of buffers, local to the user code, to the host. The host finds this information in the USEROUT[16] buffer described in the main section of this document. The calling of **UserBegin** is not maskable. All other functions that are part of the user interface will be called only if bit 0 of MODCSRB is set at the time.

When a run starts, the DSP executes a run start initialization during which it will call **UserRunInit**. It may be used to prepare data for the event procesing routines.

When events are processed by the DSP code it may call user code in two different instances. Events are processed one channel at the time. For each channel with data, **UserChannel** is called at the end of the processing, but before the energy is histogrammed. UserChannel has access to the energy, the acquired wave form (the trace) and is permitted one return value. This is the routine in which custom pulse shape analysis will be performed.

After the entire event, consisting of data from one to four channels, has been processed the function **UserEvent** may be called. It may be used in applications in which data have to be correlated across channels.

At the end of a run the closing routine may call **UserRunFinish**, typically for updating statistics and similar run end tasks.

The above mentioned routines are described below, including the interface variables and the permissible use of resources.

6.4 The interface

The interface consists of five routines and a number of global variables. Data exchange with the host computer is achieved via two data arrays that are part of the I/O parameter blocks visible to the host.

The total amount of memory available to the user comprises 2048 instructions and 1000 data words.

Host interface as supported by the DGF4C-Viewer:

UserIn[16] 16 words of input data
UserOut[16] 16 words of output data

Interface DSP routines:

UserBegin:
This routine is called after rebooting the DSP. Its purpose is to establish values for variables that need to be known before the first run may start. Address pointers to data buffers established by the user are an example. The host will need know where to write essential data to before starting a run.

Since the DSP program comes up in a default state after rebooting UserBegin will always be called. This is different for the routines listed below, which will only be called if for at least one channel bit 0 of ChannelCSRB has been set.

UserRunInit:
This function is called at each run start, for new runs as well as for resumed runs. The purpose is to precompute often needed variables and pointers here and make them available to the routines that are being called on an event-by-event basis. The variables in question would be those that depend on settings that may change in between runs.

UserChannel:
This function is called for every event and every DGF4C channel for which data are reported and for which bit 0 of the channel CSR_B (ChannelCSRB variable) has been set. It is called after all regular event processing for this channel has finished, but before the energy has been histogrammed.

UserEvent:
This function is called after all event processing for this particular event has finished. It may be used as an event finish routine, or for purposes where the event as a whole is to be examined.

UserRunFinish:
This routine is called after the run has ended, but before the host computer is notified of that fact. Its purpose is to update run summary information.

Global variables:

UserIn[16] 16 words of input data, also visible to host
UserOut[16] 16 words of output data, also visible to host

When entering UserChannel the following globals have been set by the DSP:

Atstart	Address of 1st word of the ADC trace
Tlen	Length of the ADC trace
Energy	Pulse height of the event
ChanNum	Current channel number
GSLTtimeA	high word, middle, low word
GSLTtimeB	of the 40MHz timer
GSLTtimeC	GSLT arrival time
RUNTASK	RUNTASK of the current run
E0L	lagging Energy filter, low word
E0H	lagging Energy filter, high word
E1L	leading Energy filter, low word
E1H	leading Energy filter, high word

Your return value is UretVal. It is an array of 6 words. If bit 1 of ChanCSRB is 0, only the first word is incorporated into the output data stream by the main code. See tables 4-6 in the user manual for the output data structure. If the bit is 1, up to six values are incorporated, overwriting the XIA PSA value, the USER PSA value, the GSLT time, and the reserved word in the channel header. If the run type compresses the standard nine channel header words, the number of user

return values is reduced accordingly (i.e only 2 words are available in RunTask 0x102 or 0x202, and no words in RunTask 0x103 or 0x203).

When entering UserChannel or UserEvent the address register I5 will point to the start of the current event.

Register usage:

The user routines may use all computational registers without having to restore them. However, the secondary register set cannot be used, because the XIA interrupt routines use these.

The usage of the address registers I0..I7 and the associate registers M0..M7, and L0..L7 is subject to restrictions. These are listed below for the various routines.

The associate registers L,M are preset and guaranteed as follows:

L0..L7 = 0
M0 = 0; M1 = 1; M2 = -1;
M4 = 0; M5 = 1; M6 = -1;
M3 and M7 have no guaranteed values.

UserBegin, UserRunInit, and UserRunFinish:
No further restrictions, but user code must leave the associated registers listed above in exactly this state when exiting.

UserChannel:

I5,I6,I7 L5,L6, M0,M1,M2,M4,M5,M6	These registers may not even temporarily be overwritten, because there are interrupt functions, which depend on the contents of these registers.
I0,I1,I3,I4 L0,L1,L2,L3,L4,L7	These registers may be altered, but must be restored on exit.
I2 M3,M7	These registers may be altered and need not be restored

UserEvent:

I5,I6,I7 L5,L6, M0,M1,M2,M4,M5,M6	These registers may not even temporarily be overwritten, because there are interrupt functions, which depend on the contents of these registers.
I4 L0,L1,L2,L3,L4,L7	These registers may be altered, but must be restored on exit.
I0,I1,I2,I3 M3,M7	These registers may be altered and need not be restored

6.5 Debugging tools

Besides the debugging tools that are accessible through the DGF-4C Viewer, it is also possible to download data into the DGF data buffers and call the event processing routine. This allows for an in-situ test of the newly written code and allows exploring the valid parameter space systematically or through a Monte Carlo from the host computer. For this to work the module has to halt the background activity of continuous base line measuring. Next, data have to be downloaded and the event processing started. When done the host can read the results from the known address.

The process is fairly simple. The host writes the length of the data block that is to be downloaded into the variable XDATLENGTH. Then the data are written to the linear output buffer, the address and length of which are given in the variables AOUTBUFFER and LOUTBUFFER. Next the user starts a data run, and reads the results after the run has ended.